



7007-10
(with 90 mm square semi-serrated anvil)

SPECIFICATIONS

Metric

Code No.	Clamping range on column (mm)	Stem hole (mm)	Remarks
7007-10	0 to 90	ø8, ø0.375 in	With square anvil

Note 1: Perpendicularity of the stem hole to the anvil is better than 0.4/100.

Note 2: Take note that when mounting high-accuracy Linear Gages (with resolution of 0.1 µm or better) to these stands, accuracy may be affected depending on the perpendicularity of the mounting hole to the top surface of the anvil (cosine effect).

Note 3: Compact dial indicators (bezel ø31, ø36) are not suitable for use with these stands.

DIMENSIONS

